



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

November 2, 2016

Encore Medical, L.P.
Ms. Teffany Hutto
Manager, Regulatory Affairs
9800 Metric Boulevard
Austin, Texas 7858-5445

Re: K111061

Trade/Device Name: Reverse® Shoulder Monoblock Prosthesis
Regulation Number: 21 CFR 888.3660
Regulation Name: Shoulder joint metal/polymer semi-constrained cemented prosthesis
Regulatory Class: Class II
Product Code: PHX, KWS
Dated: June 20, 2011
Received: June 21, 2011

Dear Ms. Hutto:

This letter corrects our substantially equivalent letter of July 20, 2011.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins -S

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

510(k) Number (if known): _____

K111061

Device Name: Reverse Shoulder Monoblock Prosthesis

Indications for Use:

**Reverse® Shoulder Monoblock Prosthesis
Indications for Use**

The Reverse Shoulder Monoblock Prosthesis is indicated for patients with a functional deltoid muscle with a grossly deficient rotator cuff shoulder joint with severe arthropathy or a previously failed joint replacement with a grossly deficient rotator cuff shoulder joint:

- In cases of fracture of glenohumeral joint from trauma or pathologic conditions of the shoulder, including humeral head fracture or displaced 3-or 4-part fractures of proximal humerus.
- In cases of bone defect in proximal humerus.

The patient's joint must be anatomically and structurally suited to receive the selected implant(s).

The glenoid baseplate is intended for cementless application with the addition of screws for fixation. The humeral stem is intended for cemented use only.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


(Division Sign-Off)

Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K111061

JUL 20 2011

Summary of Safety and Effectiveness

K111061

Date: July 15, 2011

Manufacturer:

DJO Surgical (legally Encore Medical, L.P.)
9800 Metric Blvd
Austin, TX 78758

Contact Person:

Teffany Hutto
Manager, Regulatory Affairs
Phone: (512) 834-6255
Fax: (512) 834-6313
Email: teffany.hutto@djosurgical.com

Product	510(k) Number and Classification	Product Code
Reverse® Shoulder Monoblock Prosthesis	K111061, Class II	KWS

Product Code	Regulation and Classification Name
KWS	Shoulder joint metal/polymer semi-constrained prosthesis per 21 CFR 888.3660

Description: This is a modification to the RSP Monoblock Stem cleared under K100741. The cleared RSP Monoblock Prosthesis system consists of a monoblock humeral stem with socket, humeral socket inserts, glenoid head, and glenoid head baseplate.

This change consists of the addition of an 8mm spacer and screw placed between the socket insert and the humeral shell portion of the RSP Monoblock stem to allow for 8mm humeral offset.

There is no change to the intended use or fundamental scientific technology. This includes no changes to packaging or sterilization.

Indications for Use: The Reverse Shoulder Monoblock Prosthesis is indicated for patients with a functional deltoid muscle with a grossly deficient rotator cuff shoulder joint with severe arthropathy or a previously failed joint replacement with a grossly deficient rotator cuff shoulder joint:

- In cases of fracture of glenohumeral joint from trauma or pathologic conditions of the shoulder, including humeral head fracture or displaced 3-or 4-part fractures of proximal humerus.
- In cases of bone defect in proximal humerus.

The patient's joint must be anatomically and structurally suited to receive the selected implant(s).

The glenoid baseplate is intended for cementless application with the addition of screws for fixation. The humeral stem is intended for cemented use only.

Predicate Device:

- DJO Surgical Reverse Shoulder Prosthesis – K041066, K051075, K100741

Comparable Features to Predicate Device(s): Features comparable to predicate devices include the same design features, materials, indications, sterilization, and intended use.

Non-Clinical Testing: Mechanical testing has demonstrated the device's ability to perform under expected conditions. Testing includes torque, lever out strength, load cycle, and screw torque.

Clinical Testing: None provided.